



# State Hygienic Laboratory

## The University of Iowa

JON MARTENS  
ATLANTIC MUNICIPAL UTILITIES  
15 W 3RD ST  
ATLANTIC, IA 50022-

Accession Number	12033
Date Sample Finalized	2012-01-19 11:20
Date Received	2012-01-17 09:30
Sample Source	Drinking Water
Project	
Date Collected	2012-01-16 10:15
Collection Site	well #6
Collection Town	ATLANTIC
Sample Description	
Client Reference	
Collector	parrott troy
Phone	712/243-1395

*Note:* Upon arrival, sample met container and preservation requirements for the analysis requested. Please review carefully your sample results for additional analyte comments or method exceptions.

### Results of Analyses

#### GCMS Volatiles, EPA 524.2

Units	mg/L
Date Analyzed	2012-01-17 16:42
Analyst	LL, JM

Analyzed In	Iowa City
Date Verified	2012-01-19 11:20
Verifier	JM

Analyte	Result	Quant Limit	MCL
Tetrachloroethylene	0.0062	0.0005	0.005

*Note:* The MCL (maximum contaminant level) is only applicable to compliance monitoring samples under the Safe Drinking Water Act (SDWA).

#### Description of Units used within this report

mg/L = Milligrams per Liter

The result(s) of this report relate only to the items analyzed. This report shall not be reproduced except in full without the written approval of the laboratory.

If you have any questions, please call Client Services at 800/421-IOWA (4692) or 319/335-4500. Thank you.

A7ZQ

40479925

1.0



Superfund

0000